



OIML Member State
Japan

OIML Certificate No.
R76/2006-JP1-17.01

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9, Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Ryoji Chubachi, President of AIST

Applicant

Name: A&D Company, Limited
Address: 3-23-14 Higashi-ikebukuro, Toshima-ku, Tokyo 170-0013, JAPAN

Manufacturer of the certified type

Name: A&D SCALES CO., LTD.
Address: 191, Inseok-ro, Deoksan-myeon, Jincheon-gun, Chungcheongbuk-do, 27856, KOREA

Identification of the certified type

Non-automatic weighing instruments
Type: HV-C/HV-CP series
Further characteristics see page 3

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, Edition 2006, for accuracy class: **III**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State
Japan



OIML Certificate No.
R76/2006-JP1-17.01

The conformity was established by the results of tests and examinations provided in the associated:

Evaluation Report N° 29-002 that includes the following test reports.

- N° 17-01-A / R76-06 that includes 40 pages
- N° 17-01-B / R76-06 that includes 20 pages
- N° 17-01-C / R76-06 that includes 20 pages
- N° 17-01-D / R76-06 that includes 26 pages
- N° 17-01-E / R76-06 that includes 10 pages

The Issuing Authority
NMIJ/AIST

Dr. R. Chubachi
President of AIST
2017-08-29



The OIML Member

Dr. Y. Miki

2017-08-29



OIML Member State
Japan

OIML Certificate No.
R76/2006-JP1-17.01

Characteristics:

Type	HV-15KC	HV-15KCP	HV-60KC	HV-60KCP	HV-200KC	HV-200KCP
Class	III					
Max	3 / 6 / 15 kg		15 / 30 / 60 kg		60 / 150 / 220 kg	
e	0.001 / 0.002 / 0.005 kg		0.005 / 0.01 / 0.02 kg		0.02 / 0.05 / 0.1 kg	
n	3000 / 3000 / 3000		3000 / 3000 / 3000		3000 / 3000 / 2200	
Min	0.02 kg		0.1 kg		0.4 kg	
Maximum Tare-balancing	15 kg		60 kg		220 kg	
Preset tare	3 kg		15 kg		60 kg	
Temperature range	- 10 °C / +40 °C					
Power supply	AC adapter and /or Dry battery (HV-**KC Type only)					
Software identification	Version number: P-1.xx (xx = 00 to 99 and shows the non-legally relevant part)					

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Evaluation Report is not permitted, although either may be reproduced in full.