



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N°R76/1992-NL-96.09

Project number 10045003

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: A.J. Nederlof

Applicant

Name: A&D Instruments
Address: Abingdon Science Park
Abingdon, Oxford OX14 3YS
United Kingdom

Manufacturer of the certified pattern

Name: A&D Instruments
Address: Abingdon Science Park
Abingdon, Oxford OX14 3YS
United Kingdom

Identification of the certified pattern

Type: HV.. and HW..
 $30 \text{ kg} \leq \text{Max} \leq 150 \text{ kg}$
 $e \geq 5 \text{ g}$
 $n \leq 7500 \text{ divisions}$

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test reports, the EC type-approval certificate and the description with number T2805 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
edition 1992
for accuracy class (III) and Class (III)

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht (NL)
Telephone +31 786332332
Telefax +31 786332309
E-mail NMI@NMI.nl

Nederlands Meetinstituut B.V. (Registered at the Chamber of Commerce Delft number 28701)

Subsidiary companies:
NMI Certin B.V. (33418)
NMI Van Swinden Laboratorium B.V. (28703)
NMI Inspecties en Kansspeltechniek B.V. (28700)
NMI International B.V. (39176)

This certificate is issued under the provision that Nederlands Meetinstituut N.V. nor its subsidiary companies accept any liability.

Reproduction of the complete certificate is allowed. Parts of the certificate may only be reproduced after written permission.



**QUALIFIED
BY STERLAB**
Reg. nr. L 029



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N°R76/1992-NL-96.09
Project number 10045003
Page 2 of 2

The conformity was established by tests described in the associated test reports:
N° R76/1992-NL-96.09A, that includes 44 pages;
N° R76/1992-NL-96.09B, that includes 41 pages.

The issuing authority
A.J. Nederlof

14 June 1996

The OIML member
G.J. Faber

14 June 1996

*
* *

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.