

Member State The Netherlands

OIML Certificate N° R76/1992-NL1-03.08

Project number 206345

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

NMi Certin B.V. Name:

Hugo de Grootplein 1, Dordrecht Address:

Person responsible: P.P.M. van Enckevort

Applicant

A&D Instruments Ltd. Name:

24, Blacklands Way Address:

Abingdon Business Park, Abingdon

Oxford OX14 1DY United Kingdom

Manufacturer of the certified pattern

A&D Company Ltd. Name:

3-23-14 Higashi-Ikebukuro Address:

> Toshima-Ku Tokyo 170 Japan

Identification of the certified pattern

Type: AD-1

n < 10000 for class (III) instruments or n < 1000 for class (III) instruments.

Temperature range -5 °C / +40 °C

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the type-approval certificate and the description with number T2178 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

edition 1992

for accuracy class (III) and (IIII)

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

Nederlands Meetinstituut Hugo de Grootplein 1 3314 EG Dordrecht

Telephone +31 78 6332332 Telefax +31 78 6332309

NMi B.V.

(Chamber of Commerce no.27.228.701)

Subsidiary companies:

NMi Van Swinden Laboratorium B.V. (27228703) NMi Certin B.V. (27.233.418) Verispect B.V. (27.228.700)

This document is issued under the provision that NMi. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



Member State
The Netherlands

OIML Certificate N° R76/1992-NL1-03.08

Project number 206345 Page 2 of 2

The conformity was established by tests described in the associated test report:

- N° R76/1992-NL1-03.08, that includes 45 pages (Indicator Ad-4402);
- N° R60/2000-NL-02.15, that includes 46 pages (Load cell LCC11).

The issuing authority P.P.M. van Enckevort Manager Certification Delft

12 March 2003

The CIML member G.J. Faber

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.