

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N°
R76/1992-NL-95.05
Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI IJkwezen B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: M. Charité

Applicant

Name: A&D Instruments Ltd
Address: Abingdon Science Park
Abingdon, Oxford
OX14 3YS United Kingdom

Manufacturer of the certified pattern

Name: A&D Instruments Ltd
Address: Abingdon Science Park
Abingdon, Oxford
OX14 3YS United Kingdom

Identification of the certified pattern

Type FS-KL
Maximum weighing capacity $6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
 $T \leq - \text{Max}$
 $n \leq 3\,000$ divisions (per partial weighing range) for class **III**
With a maximum of two partial weighing ranges
 $n \leq 1\,000$ divisions for class **III**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the EC type-approval certificate and the description with number T2290 and the documentation folder number T2290-3) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

R76

edition 1992

for accuracy class **III** or **III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht (NL)
Telephone +31 78 33 23 32
Telefax +31 78 33 23 09

Nederlands Meetinstituut N.V. (Registered at the Chamber of Commerce Delft number 28701)


Subsidiary companies:
NMI Certin B.V. (33418)
NMI Van Swinden Laboratorium B.V. (28703)
NMI IJkwezen B.V. (28700)
NMI Test- en Adviescentrum (TAC) B.V. (28702)

This certificate is issued under the provision that Nederlands Meetinstituut N.V. nor its subsidiary companies accept any liability.

Reproduction of the complete certificate is allowed. Parts of the certificate may only be reproduced after written permission.



**QUALIFIED
BY STERLAB**
Reg. nr. L 029



Nederlands Meetinstituut

OIML Certificate N°
R76/1992-NL-95.05

The conformity was established by tests described in the associated test report N° R76/1992-NL-95.05, that includes 56 pages.

The issuing authority
M. Charité


20 February 1995

The OIML member
G.J. Faber


20 February 1995

*
* *

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.

Page 2. This certificate includes 2 pages.