

OIML Certificate N°R76/1992-NL-98.20 Project number 10086223 Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name:

NMi Certin B.V.

Address.

Hugo de Grootplein 1, Dordrecht

Person responsible:

M. Charité

Applicant

Name:

A&D Instruments Ltd.

Address.

Abingdon Science Park

Abingdon, Oxford OX14 3YS

United Kingdom

Manufacturer of the certified pattern

Name:

A&D Instruments Ltd.

Address:

Abingdon Science Park

Abingdon, Oxford OX14 3YS

United Kingdom

Identification of the certified pattern

Type: SF and SG

6 kg ≤ Max ≤ 30 kg

 $e \ge 2g$

n ≤ 3000 divisions

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the type-approval certificate and the description with number T5328 and the appertaining documentation folder, with the requirements of the following Recommendation of the International Organisation of Legal Metrology (OIML):

R76

edition 1992

for accuracy class (III)



This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

Nederlands Meetinsituut Hugo de Grootplein 1 3314 EG Dordrecht Telephone +31 78 6332332

Telefax +31 78 6332309

NMi B V. (Chamber of Commerce Haaglanden

Subsidiary companies: NMi Certin B.V. (27233418) NMi Van Swinden Laboratori m B.V. (27228703) NMi International B.V. (27239176)

This document is issued under the provision that NMi. B.V. nor its subsidiary companies accept any liability

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission





Member State
The Netherlands

OIML Certificate N°R76/1992-NL-98.20

Project number 10086223

Page 2 of 2

The conformity was established by tests described in the associated test report: N° R76/1992-NL-98.20, that includes 55 pages.

The issuing authority

M. Charité

14 August 1998

The CIML member

G.J. Faber

14 August 1998

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.